

شهادة اعتماد رقم (٢٠٥٢٣ أ)



جمهورية مصر العربية
المجلس الوطني للاعتماد (إيكاك)

يشهد بأن

معامل ميراكو للاختبار والمعايرة

شركة مصر لصناعة التبريد والتكييف (ميراكو)

الكيلو (٢٨) طريق مصر اسكندرية الصحراوي - أبو رواش - المنطقة الصناعية

محافظه الجيزة - مصر

قد حصل على اعتماد المجلس حيث تم تقييمه طبقاً للمواصفة الدولية

ISO/IEC 17025:2017

في بعض المعايير في مجالات القياسات الحرارية والضغط والأبعاد

الهندسية والعزوم وبعض اختبارات أجهزة تكييف الهواء

الموضحة بالبيان المرفق رقم (٢٠٥٢٣ ب)

رقم إصدار المجال (٠٥)

إصدار رقم (٠٥) : ٠٧ أكتوبر ٢٠٢١ صالحة حتى : ٢٢ أغسطس ٢٠٢٢

شرط الحفاظ على التوافق مع المواصفات أعلاه والمتطلبات الخاصة بالمجلس

المعامل لها الحق في إصدار تقارير/شهادات تحت مظلة المجلس الوطني للاعتماد طبقاً لمجال الاعتماد المرفق

والمجلس الوطني للاعتماد موقع على اتفاقية الاعتراف المتبادل مع منظمة التعاون الدولي لاعتماد المعامل (ILAC)

في مجال اعتماد معامل المعايرة والاختبار والتحليل الطبية ومقدمي خدمة اختبارات الكفاءة الفنية وجهات التفتيش

تاريخ الاعتماد الأول: ٢٣ أغسطس ٢٠١٠

نيقين جامع

نيقين جامع

رئيس المجلس

وزير التجارة والصناعة



مهندس / هانى الدسوقي

هانى الدسوقي

المدير التنفيذي للمجلس



Accreditation Certificate No. (20523A)

**Arab Republic of Egypt
Egyptian Accreditation Council (EGAC)**

Certifies that

**Miraco Laboratories for Testing and Calibration
Misr Refrigeration Air Conditioning Company (Miraco)
Kilo (28) - Cairo Alexandria Desert Road - Abu Rawash - Industrial Zone
Giza Governorate - Egypt
ISO/IEC 17025:2017**

**Has been accredited by EGAC in compliance with the requirements of
In Some Testing of Air Conditioning Equipment and In Some Calibration
of Temperature Measurement, Pressure, Dimension and Torque**

The scope of accreditation is described in the attached schedule No. (20523B)

Scope Issue No.(05)

Issue No. (05): October 07, 2021

Valid to: August 22, 2022

**Subject to continued compliance to the above standard
and the requirements of EGAC**

**The Laboratories are accredited to issue reports/certificates under EGAC
Accreditation According to the attached scope of accreditation**

**EGAC is an ILAC MRA Signatory in the Field of Calibration, Testing, Medical Labs,
Proficiency Testing Providers and Inspection Bodies Accreditation**

1st Accreditation Date: August 23, 2010

Eng. Hany El Desouki

Hany El Desouki

EGAC Executive Director



Nevein Gamea

.....Nevein Gamea.....

**Chairman of EGAC
Minister of Trade and Industry**

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To
Miraco Laboratories for Testing and Calibration
Misr Refrigeration Air Conditioning Company (MirCO)

Kilo 28, Cairo Alexandria Desert Road, Abo Rawash, Industrial Zone, , Giza Governorate , Egypt

Schedule No.: 20523B

1st accreditation date: August 23, 2010

Issue No(05): October 7, 2021

Revision No.(-) : -----

Valid to : August 22, 2022

Materials / Products Tested	Type of Tests / Properties Measured / Range of Measurements	Standard Specifications / Techniques Used
Air Conditioning Units (Window, Split, Package)	Determination of Sound Power and Sound Pressure Levels at frequency 125 Hz to 10 kHz By Using Precision Method, Acoustics	ISO 3741:2010 - Reference Sound Source S.N 1818653 - Real Time Frequency Analyzer S.N 2684302 - Microphone S.N 1784034, 1811692, 1811694 - Microphone Preamplifier S.N 1834741, 1828407, 1828408 - Sound Level Calibrator S.N 1838979, 3008703 - Resistance Temperature Detector S.N 0019, 0020

Asman

M. El-Maadi

Kornish El-Maadi, Riad El-Maadi Tower 1 - Cairo - Egypt

Tel.: (202) 25275220/5/6/7

Fax: (202) 25275224

F4WI4TCL
1 / Dec 2018

Page 7 of 7

كورنيش المعادى - برج رياض المعادى ١ - القاهرة - مصر

تليفون : ٢٥٢٧٥٢٢٠/٥/٦/٧ (٢٠٢)

فاكس : ٢٥٢٧٥٢٢٤ (٢٠٢)